



# **EXPRESS qPCR SuperMixes and Two-Step qRT-PCR Kits**

**Catalog nos. 11783-200/01K, 11785-200/01K,  
11793-200/01K, and 11795-200/01K**

**Rev. Date: 28 June 2010**  
Manual part no. A10329

MAN0000691

**User Manual**



# Table of Contents

Kit Contents and Storage .....	iv
Overview .....	1
Instrument Compatibility .....	4
<b>Methods .....</b>	<b>5</b>
First-Strand cDNA Synthesis .....	5
General qPCR Guidelines and Parameters.....	9
Universal Kits—Guidelines and Protocols.....	11
Kits with Premixed ROX—Guidelines and Protocols.....	15
Troubleshooting .....	17
<b>Appendix .....</b>	<b>19</b>
Additional Products .....	19
Technical Support .....	20
Purchaser Notification .....	22
References .....	24

## Kit Contents and Storage

---

### Kit Components and Storage

EXPRESS qPCR SuperMixes and Two-Step qRT-PCR Kits are shipped on dry ice. The components in each kit are listed below.

**Storage:** Store all components at  $-20^{\circ}\text{C}$  for long-term storage. EXPRESS qPCR SuperMixes may be stored at  $4-8^{\circ}\text{C}$  for up to one month.

<b>EXPRESS qPCR Supermix Universal</b>	<b>11785-200</b>	<b>11785-01K</b>
EXPRESS qPCR SuperMix Universal	5 ml	5 × 5 ml
ROX Reference Dye	500 $\mu\text{l}$	5 × 500 $\mu\text{l}$

<b>EXPRESS qPCR Supermix with Premixed ROX</b>	<b>11795-200</b>	<b>11795-01K</b>
EXPRESS qPCR SuperMix with Premixed ROX	5 ml	5 × 5 ml

<b>EXPRESS Two-Step Superscript<sup>®</sup> qRT-PCR Universal</b>	<b>11783-200</b>	<b>11783-01K</b>
EXPRESS qPCR SuperMix Universal	5 ml	5 × 5 ml
ROX Reference Dye	500 $\mu\text{l}$	5 × 500 $\mu\text{l}$
SuperScript <sup>®</sup> VILO <sup>™</sup> cDNA Synthesis Kit – 5X VILO <sup>™</sup> Reaction Mix – 10X SuperScript <sup>®</sup> Enzyme Mix	50 rxns (20 $\mu\text{l}$ each)	250 rxns (20 $\mu\text{l}$ each)

<b>EXPRESS Two-Step Superscript<sup>®</sup> qRT-PCR with Premixed ROX</b>	<b>11793-200</b>	<b>11793-01K</b>
EXPRESS qPCR SuperMix with Premixed ROX	5 ml	5 × 5 ml
SuperScript <sup>®</sup> VILO <sup>™</sup> cDNA Synthesis Kit – 5X VILO <sup>™</sup> Reaction Mix – 10X SuperScript <sup>®</sup> Enzyme Mix	50 rxns (20 $\mu\text{l}$ each)	250 rxns (20 $\mu\text{l}$ each)

### Product Qualification

The Certificate of Analysis (CofA) provides detailed quality control information for each product. The CofA is available on our website at [www.invitrogen.com/cofa](http://www.invitrogen.com/cofa), and is searchable by product lot number, which is printed on each box.

---

# Overview

---

## Introduction

EXPRESS qPCR SuperMixes and Two-Step qRT-PCR Kits provide components for real-time quantitative PCR (qPCR) and two-step reverse-transcription qPCR (qRT-PCR). Components are provided in convenient SuperMix formats and are compatible with both rapid and standard qPCR cycling conditions.

All EXPRESS qPCR SuperMixes include Platinum® *Taq* DNA polymerase, MgCl<sub>2</sub>, heat-labile uracil DNA glycosylase (UDG), dNTPs (with dUTP instead of dTTP), and stabilizers. These mixes have been formulated for use with fluorogenic probe-based technology (*e.g.*, TaqMan® probes) or fluorogenic primers (*e.g.*, LUX™ Primers).

- **SuperMix with Premixed ROX:** The qPCR SuperMix with premixed ROX includes ROX Reference Dye at a final concentration of 500 nM to normalize the fluorescent signal on instruments that are compatible with this option.
- **Universal SuperMix:** The Universal SuperMix includes ROX as a separate component for instruments that use ROX at a different concentration or do not require ROX.
- **Two-Step qRT-PCR Kits:** These kits include a qPCR SuperMix and the SuperScript® VILO™ cDNA Synthesis Kit for cDNA synthesis prior to qPCR. The VILO™ kit provides enhanced cDNA synthesis efficiency and can be used with very low and very high amounts of input RNA (from 1 pg up to 2.5 µg total RNA in a 20-µl reaction), giving a linear response in message abundance as measured by qPCR

---

*Continued on next page*

# Overview, continued

---

## Advantages of the Kits

- **Highly robust qPCR SuperMix** can accommodate a wide range of cycling conditions and reaction volumes, and combines highly sensitive detection with a broad quantification range
  - **Platinum® Taq DNA Polymerase** provides an automatic “hot start” in PCR for increased sensitivity, specificity, and yield, and has a short activation time for the rapid cycling of fast qPCR instruments
  - **Heat-labile UDG and dUTP** in the SuperMix prevent amplification of carryover PCR products between reactions, and the heat-labile form of the enzyme is completely inactivated during normal qPCR cycling, eliminating any downstream degradation of amplicons
  - **SuperScript® VILO™ cDNA Synthesis Kit**, included in the two-step kits, provides high yields of cDNA and linear output over a very broad range of RNA input quantities
- 

## Platinum® Taq DNA Polymerase

Platinum® Taq DNA Polymerase is recombinant Taq DNA polymerase complexed with proprietary antibodies that block polymerase activity at ambient temperatures (Chou *et al.*, 1992; Sharkey *et al.*, 1994). Activity is restored after the initial denaturation step in PCR cycling, providing an automatic hot start in qPCR for increased sensitivity, specificity, and yield.

---

## Uracil DNA Glycosylase (UDG)

Heat-labile UDG and dUTP in the qPCR SuperMix prevent the reamplification of carryover PCR products between reactions (Lindahl *et al.*, 1977; Longo *et al.*, 1990). dUTP ensures that any amplified DNA will contain uracil, while heat-labile UDG removes uracil residues from single- or double-stranded DNA.

The heat-labile form of UDG used in this kit is completely inactivated at temperatures of 50°C and higher and will not degrade amplicons following qPCR, thus enabling their use for downstream applications such as cloning.

---

*Continued on next page*

## Overview, continued

---

### **ROX Reference Dye**

ROX Reference Dye is either premixed in the SuperMix or included as a separate component to normalize the fluorescent signal between reactions for instruments that are compatible with this option.

---

### **Additional Materials Required**

The following items are supplied by the user:

- Template DNA
  - Gene-specific fluorescent primers or primer/probe combinations
  - DEPC-treated water
  - Microcentrifuge
  - Thermal cycler
  - Optional: Normalization dye for instruments that do not use ROX
  - PCR tubes/plates
-

# Instrument Compatibility

---

## Universal SuperMix

EXPRESS qPCR SuperMix Universal includes ROX Reference Dye as a separate tube, and can be used with a wide range of real-time instruments, including the following:

- **Applied Biosystems:** 7900HT, 7300, 7500, StepOne™, StepOnePlus™, GeneAmp® 5700, and PRISM® 7000 and 7700
  - **Bio-Rad/MJ Research:** iCycler® iQ, iQ5, and MyiQ™; DNA Engine Opticon® and Opticon® 2; and Chromo4™ Real-Time Detector
  - **Cepheid:** Smart Cycler®
  - **Corbett Research:** Rotor-Gene™ 3000
  - **Eppendorf:** Mastercycler® ep *realplex*
  - **Roche:** LightCycler® 480
  - **Stratagene:** Mx3000P™, Mx3005P™, and Mx4000®
- 

## Kits with Premixed ROX

EXPRESS qPCR SuperMix with Premixed ROX can be used with real-time instruments that are compatible with ROX Reference Dye at a final concentration of 500 nM. These include the following **Applied Biosystems** instruments:

- 7900HT
  - 7300
  - StepOne™
  - StepOnePlus™
  - GeneAmp® 5700
  - PRISM® 7000 and 7700
-



# Methods

## First-Strand cDNA Synthesis

---

### Introduction

This section provides guidelines and a protocol for first-strand cDNA synthesis from RNA using the SuperScript® VILO™ cDNA Synthesis Kit, which is included with the two-step kits and is also available separately. If you are performing qPCR using DNA from another source, you can skip this section.

---

### SuperScript® VILO™ cDNA Synthesis Kit

This cDNA synthesis kit is included with the EXPRESS Two-Step SuperScript® qRT-PCR Kits, and is also available separately (see page 19 for ordering information). It provides enhanced cDNA synthesis efficiency and can be used with very low and very high amounts of input RNA (up to 2.5 µg total RNA in a 20-µl reaction), giving a linear response in message abundance as measured by qPCR.

The **10X SuperScript® Enzyme Mix** includes SuperScript® III RT, RNaseOUT™ Recombinant Ribonuclease Inhibitor, and a proprietary helper protein.

The **5X VILO™ Reaction Mix** includes random primers, MgCl<sub>2</sub>, and dNTPs in a buffer formulation that has been optimized for qRT-PCR.

---

### Input RNA

- Starting material using the VILO™ kit can range up to 2.5 µg total RNA in a 20-µl cDNA synthesis reaction.
  - To isolate total RNA, we recommend TaqMan® Gene Expression Assays, TRIzol® Reagent, or the PureLink™ 96 Total RNA Purification Kit (see page 19). Isolation of mRNA is typically not necessary, although incorporating this step may improve the yield of specific cDNAs.
  - High-quality, intact RNA is essential for accurate quantification in qRT-PCR.
  - DNase I, Amplification Grade, may be used to eliminate genomic DNA contamination from the total RNA (see page 19).
- 

*Continued on next page*

# First-Strand cDNA Synthesis, continued

---

## General Handling of RNA

When working with RNA:

- Use proper microbiological aseptic technique.
- Wear latex gloves while handling reagents, materials, and RNA samples to prevent RNase contamination.
- Use disposable, individually wrapped, sterile plasticware for all procedures.
- Use aerosol-resistant pipette tips.
- Dedicate a separate set of pipettes, buffers, and enzymes for RNA work.
- Use RNase-free microcentrifuge tubes. To decontaminate untreated tubes, soak overnight in a 0.01% (v/v) aqueous solution of diethylpyrocarbonate (DEPC), rinse with sterile distilled water, and autoclave.

RNase *Away*<sup>™</sup> Reagent, a non-toxic solution available from Invitrogen, can be used to remove RNase contamination from surfaces.

---

## Determining Total RNA Quality

Total RNA quality can be analyzed using a bioanalyzer such as the Agilent 2100 bioanalyzer with an RNA LabChip<sup>®</sup>. Alternatively, total RNA can be analyzed by agarose gel electrophoresis. RNA isolated using the PureLink<sup>™</sup> kits or TRIzol<sup>®</sup> Reagent typically has a 28S-to-18S band ratio of >1.5. RNA is judged to be intact if discreet 28S and 18S ribosomal RNA bands are observed.

---

*Continued on next page*

# First-Strand cDNA Synthesis, continued

---

## Determining Total RNA Yield

Total RNA can be quantitated using the Quant-iT™ RNA Assay Kit or UV absorbance at 260 nm.

### Quant-iT™ RNA Assay Kit

The Quant-iT™ RNA Assay Kit provides a rapid, sensitive, and specific method for RNA quantitation with minimal interference from DNA, protein, or other common contaminants that affect UV absorbance readings.

The kit contains a quantitation reagent and pre-diluted standards for a standard curve. The assay is performed in a microtiter plate and can be read using a standard fluorescent microplate reader.

### UV Absorbance

1. Dilute an aliquot of the total RNA sample in 10 mM Tris-HCl, pH 7.5. Mix well. Transfer to a cuvette (1-cm path length).

**Note:** The RNA must be in a neutral pH buffer to accurately measure the UV absorbance.

2. Determine the OD<sub>260</sub> of the solution using a spectrophotometer blanked against 10 mM Tris-HCl, pH 7.5.

Calculate the amount of total RNA using the following formula:

$$\text{Total RNA } (\mu\text{g}) = \text{OD}_{260} \times [40 \mu\text{g}/(1 \text{OD}_{260} \times 1 \text{ml})] \times \text{dilution factor} \times \text{total sample volume (ml)}$$

### Example:

Total RNA was eluted in water in a total volume of 150  $\mu\text{l}$ . A 40- $\mu\text{l}$  aliquot of the eluate was diluted to 500  $\mu\text{l}$  in 10 mM Tris-HCl, pH 7.5. An OD<sub>260</sub> of 0.188 was obtained. The amount of RNA in the sample is:

$$\text{Total RNA } (\mu\text{g}) = 0.188 \times [40 \mu\text{g}/(1 \text{OD}_{260} \times 1 \text{ml})] \times 12.5 \times 0.15 = 14.1 \mu\text{g}$$

---

*Continued on next page*

# First-Strand cDNA Synthesis, continued

---

## Guidelines for cDNA Synthesis

- Shorter incubation times and/or higher temperatures may be used (*e.g.*, 50°C for 30 minutes), but may result in reduced yields of cDNA.
  - For increased yields of cDNA, longer incubation times may be used (up to 120 minutes at 42°C).
- 

## cDNA Synthesis Protocol

The following protocol has been optimized for generating first-strand cDNA using the SuperScript® VILO™ cDNA Synthesis Kit. The reaction volume may be scaled as needed up to 100 µl.

1. For a single reaction, combine the following components in a tube on ice. For multiple reactions, prepare a master mix without RNA.

5X VILO™ Reaction Mix	4 µl
10X SuperScript® Enzyme Mix	2 µl
RNA (up to 2.5 µg)	x µl
DEPC-treated water	to 20 µl

2. Gently mix tube contents and incubate at 25°C for 10 minutes.
  3. Incubate tube at 42°C for 60 minutes.
  4. Terminate the reaction at 85°C at 5 minutes.
  5. Use diluted or undiluted cDNA in qPCR (see the guidelines for cDNA use in qPCR on page 9). Alternatively, store the cDNA at -20°C until use.
-

# General qPCR Guidelines and Parameters

---

## Reaction Setup and Conditions

- Starting material for qPCR can be cDNA, genomic DNA, or plasmid DNA
  - Maintain a sterile environment when handling DNA to avoid any contamination from DNases
  - Make sure all equipment that comes in contact with DNA is sterile, including pipette tips and microcentrifuge tubes
  - qPCR reaction volumes can be scaled from 5  $\mu$ l to 100  $\mu$ l, depending on the instrument.
  - For instrument-specific guidelines, see the section for each type of SuperMix.
- 

## cDNA

See pages 5–8 for a cDNA synthesis protocol using the SuperScript® VILO™ cDNA Synthesis Kit. If you are using cDNA as starting material:

- Use undiluted or diluted cDNA.
  - Up to 10% of the qPCR reaction volume may be undiluted cDNA (*e.g.*, for a 20- $\mu$ l qPCR, use up to 2  $\mu$ l of undiluted cDNA).
  - Note that detecting high-abundance genes in undiluted cDNA may result in very low CTs in qPCR, leading to reduced quantification accuracy. Prepare a dilution series of the cDNA template for the most accurate results.
- 

## Genomic or Plasmid DNA

If you are using genomic or plasmid DNA as starting material:

- Use up to 100 ng of genomic DNA or 10–10<sup>7</sup> copies of plasmid DNA in a 10- $\mu$ l volume
  - Note that 1  $\mu$ g of plasmid DNA contains 9.1  $\times$  10<sup>11</sup> copies divided by the plasmid size in kilobases
- 

*Continued on next page*

# General qPCR Guidelines and Parameters, continued

---

## Fluorescent Primers

LUX™ Primers, available separately from Invitrogen, are a fluorescent primer-based detection technology consisting of one gene-specific primer labeled with a single fluorophore and a corresponding unlabeled primer. The labeled primer is designed with the fluorophore near the 3' end in a hairpin structure that effectively quenches fluorescence prior to PCR, making a separate quenching moiety unnecessary. When the primer becomes incorporated into double-stranded PCR product, the fluorophore is de-quenched, resulting in a significant increase in fluorescent signal.

LUX™ Primers are available in pre-designed formats ([www.invitrogen.com/lux](http://www.invitrogen.com/lux)) or can be designed for specific targets using the D-LUX™ Designer ([www.invitrogen.com/dluxdesigner](http://www.invitrogen.com/dluxdesigner)).

A final concentration of 200 nM per primer is effective for most reactions. Doubling the amount of reverse primer (to 400 nM) may improve the performance of certain reactions. Optimal results may require a primer titration between 100 and 500 nM.

---

## Fluorescent Probe-Based Technologies

Fluorescent probe-based technologies, such as TaqMan® Gene Expression Assays, use two gene-specific primers with a fluorescent labeled probe that emits a signal when incorporated into the PCR product. These probe-primer combinations may be packaged as predesigned gene-specific assays or custom designed for a target of interest.

Consult the documentation provided with your fluorescent probe product to determine the optimal concentration in qPCR. EXPRESS qPCR Kits were developed using TaqMan® Gene Expression Assays provided at a 20X concentration as well as custom-designed mixtures in which the final primer concentration was 500 nM each and the final probe concentration was 200 nM.

---

## Melting Curve Analysis

Melting curve analysis may be used with fluorescent primers to identify the presence of primer dimers and analyze the specificity of the reaction. Note that melting curve analysis cannot be used with fluorescent probe-based technologies. Program for melting curve analysis using the instructions provided with your specific instrument.

---

# Universal Kits—Guidelines and Protocols

---

## Introduction

This section provides guidelines and protocols for using EXPRESS qPCR SuperMix Universal.

---

## Additional Materials Required

The following items are supplied by the user:

- DNA starting material
  - DEPC-treated water
  - Fluorescent primers/probes (see page 10 for information)
  - Microcentrifuge
  - Thermal cycler
  - PCR tubes/plates
- 

## ROX Reference Dye Concentration

ROX Reference Dye is supplied as a separate tube in the Universal kits. ROX is recommended for fluorescence normalization on Applied Biosystems instruments, and is optional for Stratagene and Eppendorf instruments. It is not required on other instruments.

ROX is composed of a glycine conjugate of 5-carboxy-X-rhodamine, succinimidyl ester and is supplied at a concentration of 25  $\mu\text{M}$ .

Use the following table to determine the amount of 25- $\mu\text{M}$  ROX to use with a particular instrument:

Instrument	Amount of ROX per 20- $\mu\text{l}$ reaction	Effective Fold Concentration of 25- $\mu\text{M}$ ROX	Final ROX Concentration
AB 7300, 7900HT, StepOne™, StepOnePlus™, and PRISM® 7000 and 7700	0.4 $\mu\text{l}$	50X	500 nM
AB 7500; Stratagene Mx3000P™, Mx3005P™, and Mx4000®	0.04 $\mu\text{l}$	500X	50 nM

---

*Continued on next page*

## Universal Kits, continued

---

### General Cycling Programs

The following cycling programs have been developed as a general starting point when using EXPRESS qPCR SuperMix Universal. The fast cycling program was developed using the AB 7500 in Fast mode.

**Note:** This mix is highly robust and can be used with a wide range of cycling programs on different instruments. If you have an alternative program that you want to use, test it with this mix.

<b>Fast Cycling Program (developed using the AB 7500 in Fast mode)</b>	<b>Standard Cycling Program</b>
95°C for 20 seconds	50°C for 2 minutes (UDG incubation)
40 cycles of:	95°C for 2 minutes
95°C for 3 seconds	40 cycles of:
60°C for 30 seconds	95°C for 15 seconds
<i>Fluorescent primers only:</i>	60°C for 1 minute
Melting curve analysis: 60°C–95°C	<i>Fluorescent primers only:</i>
(refer to instrument manual for specific programming)	Melting curve analysis: 60°C–95°C
	(refer to instrument manual for specific programming)

---

*Continued on next page*



## Universal Kits, continued

### Roche LightCycler® 480 Cycling Program

The following cycling program is specific for the Roche LightCycler® 480 with a 96-well or 384-well plate when using EXPRESS qPCR SuperMix Universal. For detailed programming instructions, consult the instrument manual.

Program Name	Cycles	Analysis Mode
Pre-incubation	1	None
Amplification	40–45	Quantification
Cooling	1	None

Target (°C)	Acquisition Mode	Hold Time (hh:mm:ss)	Ramp Rate (°C/s)	
			96-well	384-well
<b>Pre-incubation</b>				
95	None	00:05:00	4.4 (or 2.0) <sup>(1)</sup>	4.8
<b>Amplification</b>				
95	None	00:00:10	4.4 (or 2.0) <sup>(1)</sup>	4.8
Primer T <sub>m</sub> minus 5°C <sup>(2)</sup>	None	00:00:15– 00:00:50 <sup>(3)</sup>	2.2 (if ≥ 50°C) <sup>(4)</sup> or 1.5 (if < 50°C)	2.5 (if ≥ 50°C) <sup>(4)</sup> or 2.0 (if < 50°C)
72	Single	00:00:01	4.4 (or 2.0) <sup>(1)</sup>	4.8
<b>Cooling</b>				
40	None	00:00:10	1.5	2.0
<sup>(1)</sup> A ramp rate of 2.0°C/s is recommended for reaction volumes of 50 µl or greater. <sup>(2)</sup> The annealing temperature will vary depending on the melting temperature (T <sub>m</sub> ) of the primers. Use primer T <sub>m</sub> minus 5°C as a general starting point. <sup>(3)</sup> Longer annealing and extension times may result in greater precision in target quantification. <sup>(4)</sup> The ramp rate depends on the annealing temperature (see note 2 above). Use the higher ramp rate if the annealing temperature is ≥50°C.				

Continued on next page

## Universal Kits, continued

---

### 384-Well Plate Volumes

For 384-well plates, we recommend a maximum reaction volume of 10  $\mu$ l per well.

---

### qPCR Protocol

Use the protocol below as a general starting point for qPCR with EXPRESS qPCR SuperMix Universal. Scale the reaction volume as needed for your real-time instrument.

ROX is recommended for Applied Biosystems instruments and optional for Stratagene and Eppendorf instruments (see page 11).

1. Set up reactions on ice. Volumes for a 20- $\mu$ l reaction are provided; component volumes can be scaled as needed. For 384-well plates, we recommend a maximum reaction volume of 10  $\mu$ l per well. **Always prepare a master mix of common components for multiple reactions.**

	<u>20-<math>\mu</math>l rxn</u>
EXPRESS qPCR SuperMix Universal	10 $\mu$ l
Fluorescent primer/probe mix (conc. and volume specified by manufacturer)	X $\mu$ l
ROX Reference Dye (25 $\mu$ M)	0.4 $\mu$ l/0.04* $\mu$ l
Template DNA (see page 9)	X $\mu$ l
DEPC-treated water	to 20 $\mu$ l

\*See the table on page 11 for the amount/concentration of ROX to use for your specific instrument.

2. Prepare no-template control (NTC) reactions to test for DNA contamination of the enzyme/primer mixes.
  3. Cap or seal each PCR tube/plate, and gently mix. Make sure that all components are at the bottom of the tube/plate; centrifuge briefly if needed.
  4. Place reactions in a real-time instrument programmed as described on the previous pages. Collect data and analyze results.
  5. **Optional:** The specificity of the PCR products can be checked by agarose gel electrophoresis.
-

# Kits with Premixed ROX—Guidelines and Protocols

---

## Introduction

This section provides guidelines and protocols for using EXPRESS qPCR SuperMix with Premixed ROX.

---

## Additional Materials Required

The following items are supplied by the user:

- DNA starting material
  - DEPC-treated water
  - Gene-specific primers
  - Microcentrifuge
  - Thermal cycler
  - PCR tubes/plates
- 

## Premixed ROX Concentration

ROX Reference Dye is included in the SuperMix at a final concentration of 500 nM, which is compatible with Applied Biosystems 7900HT, 7300, StepOne™, StepOnePlus™, GeneAmp® 5700, and PRISM® 7000 and 7700.

---

## Cycling Programs — Kits with Premixed ROX

The following cycling programs have been developed as a general starting point when using EXPRESS qPCR SuperMix with Premixed ROX. The fast cycling program is designed for the AB 7900HT and StepOne™.

**Note:** This mix is highly robust and can be used with a wide range of cycling programs. If you have an alternative program that you want to use, test it with this mix.

<b>Fast Cycling Program (developed using the AB 7900HT and StepOne™)</b>	<b>Standard Cycling Program</b>
95°C for 20 seconds	50°C for 2 minutes (UDG incubation)
40 cycles of:	95°C for 2 minutes
95°C for 1 second	40 cycles of:
60°C for 20 seconds	95°C for 15 seconds
<i>Fluorescent primers only:</i>	60°C for 1 minute
Melting curve analysis: 60°C–95°C	<i>Fluorescent primers only:</i>
(refer to instrument manual for specific programming)	Melting curve analysis: 60°C–95°C
	(refer to instrument manual for specific programming)

---

*Continued on next page*

## Kits with Premixed ROX, continued

---

### 384-Well Plate Volumes

For 384-well plates, we recommend a maximum reaction volume of 10  $\mu$ l per well.

---

### qPCR Protocol — Kits with Premixed ROX

Use the protocol below as a general starting point for using EXPRESS qPCR SuperMix with Premixed ROX. Scale the reaction volume as needed for your real-time instrument.

1. Set up reactions on ice. Volumes for a 20- $\mu$ l reaction are provided; component volumes can be scaled as needed. For 384-well plates, we recommend a maximum reaction volume of 10  $\mu$ l per well. **Always prepare a master mix of common components for multiple reactions.**

	<u>20-<math>\mu</math>l rxn</u>
EXPRESS qPCR SuperMix with Premixed ROX	10 $\mu$ l
Fluorescent primer/probe mix (conc. and volume specified by manufacturer)	X $\mu$ l
Template DNA (see page 9)	X $\mu$ l
DEPC-treated water	to 20 $\mu$ l

2. Prepare no-template control (NTC) reactions to test for DNA contamination of the enzyme/primer mixes.
  3. Cap or seal each PCR tube/plate, and gently mix. Make sure that all components are at the bottom of the tube/plate; centrifuge briefly if needed.
  4. Place reactions in a real-time instrument programmed as described on the previous page. Collect data and analyze results.
  5. **Optional:** The specificity of the PCR products can be checked by agarose gel electrophoresis.
-

# Troubleshooting

<b>Problem</b>	<b>Cause</b>	<b>Solution</b>
Signals are present in no-template controls, and/or multiple peaks are present in the melting curve graph	Template or reagents are contaminated by nucleic acids (DNA, cDNA)	Use melting curve analysis and/or run the PCR products on a 4% agarose gel after the reaction to identify contaminants. Take standard precautions to avoid contamination when preparing your PCR reactions. Ideally, amplification reactions should be assembled in a DNA-free environment. We recommend using aerosol-resistant barrier tips.
	Primer dimers or other primer artifacts are present	Use melting curve analysis to identify primer dimers. We recommend using validated pre-designed primer/probe sets or design them using dedicated software programs or databases. Contamination or truncated or degraded primers/probes can lead to artifacts. Check the purity of your primers and probes by gel electrophoresis.
No PCR product is evident, either in the qPCR graph or on a gel	The protocol was not followed correctly	Verify that all steps have been followed and the correct reagents, dilutions, volumes, and cycling parameters have been used.
	Template contains inhibitors, nucleases, or proteases, or has otherwise been degraded.	Purify or re-purify your template.
	Primer/probe design is suboptimal	We recommend using validated pre-designed primers/probes or design them using dedicated software programs or databases.

*Continued on next page*

## Troubleshooting, continued

<b>Problem</b>	<b>Cause</b>	<b>Solution</b>
PCR product is evident on a gel, but not in the qPCR graph	qPCR instrument settings are incorrect	Confirm that you are using the correct instrument settings (dye selection, reference dye, filters, and acquisition points).
	Problems with your specific qPCR instrument	See your instrument manual for tips and troubleshooting.
PCR efficiency is above 110%	Template contains inhibitors, nucleases, or proteases, or has otherwise been degraded.	Purify or re-purify your template. Inhibitors in the template may result in changes in PCR efficiency between dilutions.
	Too much sample added to reactions	Decrease the concentration of sample; see the guidelines for sample concentration on page 9
	Nonspecific products may be amplified.	Use melting curve analysis if possible, and/or run the PCR products on a 4% agarose gel after the reaction to identify contaminants. Suboptimal primer design may lead to nonspecific products. Use validated pre-designed primers/probes or design them using dedicated software programs or primer databases.
PCR efficiency is below 90%	The PCR conditions are suboptimal	Verify that the reagents you are using have not been freeze-thawed multiple times and have not remained at room temperature for too long. Verify that the amount of primers you are using is correct.

# Appendix

## Additional Products

---

### Additional Products

Related products are available separately from Invitrogen. Ordering information is provided below. For more information, visit our website at [www.invitrogen.com](http://www.invitrogen.com) or contact Technical Service (page 19).

Product	Quantity	Catalog no.
TaqMan® Gene Expression Assays	visit <a href="http://www.invitrogen.com/tagman">www.invitrogen.com/tagman</a>	
LUX™ Fluorogenic Primers	visit <a href="http://www.invitrogen.com/lux">www.invitrogen.com/lux</a>	
SuperScript® VILO™ cDNA Synthesis Kit	50 rxns 250 rxns	11754-050 11754-250
RNase <i>Away</i> ™ Reagent	250 ml	10328-011
DNase I, Amplification Grade	100 units	18068-015
TRIzol® Reagent	100 ml 200 ml	15596-026 15596-018
PureLink™ Genomic DNA Mini Kit	50 preps 250 preps	K1820-01 K1820-02
Quant-iT™ RNA Assay Kit	1 kit	Q-33140
Quant-iT™ DNA Assay Kit, High Sensitivity	1000 assays	Q33120
Quant-iT™ DNA Assay Kit, Broad-Range	1000 assays	Q33130

---

# Technical Support

---

## Web Resources



Visit the Invitrogen website at [www.invitrogen.com](http://www.invitrogen.com) for:

- Technical resources, including manuals, vector maps and sequences, application notes, SDSs, FAQs, formulations, citations, handbooks, etc.
  - Complete technical support contact information
  - Access to the Invitrogen Online Catalog
  - Additional product information and special offers
- 

## Contact Us

For more information or technical assistance, call, write, fax, or email. Additional international offices are listed on our website ([www.invitrogen.com](http://www.invitrogen.com)).

### Corporate Headquarters:

5791 Van Allen Way  
Carlsbad, CA 92008 USA  
Tel: 1 760 603 7200  
Tel (Toll Free): 1 800 955 6288  
Fax: 1 760 602 6500  
E-mail:  
[tech\\_support@invitrogen.com](mailto:tech_support@invitrogen.com)

### Japanese Headquarters:

LOOP-X Bldg. 6F  
3-9-15, Kaigan  
Minato-ku, Tokyo 108-0022  
Tel: 81 3 5730 6509  
Fax: 81 3 5730 6519  
E-mail:  
[jpinfo@invitrogen.com](mailto:jpinfo@invitrogen.com)

### European Headquarters:

Inchinnan Business Park  
3 Fountain Drive  
Paisley PA4 9RF, UK  
Tel: +44 (0) 141 814 6100  
Tech Fax: +44 (0) 141 814 6117  
E-mail:  
[eurotech@invitrogen.com](mailto:eurotech@invitrogen.com)

---

## SDS

Safety Data Sheets (SDSs) are available on our website at [www.invitrogen.com/sds](http://www.invitrogen.com/sds).

---

## Certificate of Analysis

The Certificate of Analysis (CofA) provides detailed quality control information for each product. The CofA is available on our website at [www.invitrogen.com/cofa](http://www.invitrogen.com/cofa), and is searchable by product lot number, which is printed on each box.

---

*Continued on next page*



# Technical Support, continued

---

## Limited Warranty

Invitrogen (a part of Life Technologies Corporation) is committed to providing our customers with high-quality goods and services. Our goal is to ensure that every customer is 100% satisfied with our products and our service. If you should have any questions or concerns about an Invitrogen product or service, contact our Technical Support Representatives.

All Invitrogen products are warranted to perform according to specifications stated on the certificate of analysis. The Company will replace, free of charge, any product that does not meet those specifications. This warranty limits the Company's liability to only the price of the product. No warranty is granted for products beyond their listed expiration date. No warranty is applicable unless all product components are stored in accordance with instructions. The Company reserves the right to select the method(s) used to analyze a product unless the Company agrees to a specified method in writing prior to acceptance of the order.

Invitrogen makes every effort to ensure the accuracy of its publications, but realizes that the occasional typographical or other error is inevitable. Therefore the Company makes no warranty of any kind regarding the contents of any publications or documentation. If you discover an error in any of our publications, report it to our Technical Support Representatives.

**Life Technologies Corporation shall have no responsibility or liability for any special, incidental, indirect or consequential loss or damage whatsoever. The above limited warranty is sole and exclusive. No other warranty is made, whether expressed or implied, including any warranty of merchantability or fitness for a particular purpose.**

---

# Purchaser Notification

---

**Limited Use  
Label License  
No. 5: Invitrogen  
Technology**

The purchase of this product conveys to the buyer the non-transferable right to use the purchased amount of the product and components of the product in research conducted by the buyer (whether the buyer is an academic or for-profit entity). The buyer cannot sell or otherwise transfer (a) this product (b) its components or (c) materials made using this product or its components to a third party or otherwise use this product or its components or materials made using this product or its components for Commercial Purposes. The buyer may transfer information or materials made through the use of this product to a scientific collaborator, provided that such transfer is not for any Commercial Purpose, and that such collaborator agrees in writing (a) not to transfer such materials to any third party, and (b) to use such transferred materials and/or information solely for research and not for Commercial Purposes. Commercial Purposes means any activity by a party for consideration and may include, but is not limited to: (1) use of the product or its components in manufacturing; (2) use of the product or its components to provide a service, information, or data; (3) use of the product or its components for therapeutic, diagnostic or prophylactic purposes; or (4) resale of the product or its components, whether or not such product or its components are resold for use in research. For products that are subject to multiple limited use label licenses, the terms of the most restrictive limited use label license shall control. Life Technologies Corporation will not assert a claim against the buyer of infringement of patents owned or controlled by Life Technologies Corporation which cover this product based upon the manufacture, use or sale of a therapeutic, clinical diagnostic, vaccine or prophylactic product developed in research by the buyer in which this product or its components was employed, provided that neither this product nor any of its components was used in the manufacture of such product. If the purchaser is not willing to accept the limitations of this limited use statement, Life Technologies is willing to accept return of the product with a full refund. For information about purchasing a license to use this product or the technology embedded in it for any use other than for research use please contact Out Licensing, Life Technologies, 5791 Van Allen Way, Carlsbad, California 92008; Phone (760) 603-7200 or e-mail: [outlicensing@lifetech.com](mailto:outlicensing@lifetech.com).

---

*Continued on next page*

# Purchaser Notification, continued

---

**Limited Use  
Label License  
No. 14: Direct  
Inhibition by  
Anti-Polymerase  
Antibodies**

Licensed to Life Technologies Corporation, under U.S. Patent Nos. 5,338,671; 5,587,287; and foreign equivalents for use in research only.

---

**Limited Use  
Label License  
No. 274: 5'  
Nuclease  
Process**

A license to perform the 5' nuclease process for research requires the use of a Licensed 5' Nuclease Kit (containing Licensed Probe), or the combination of an Authorized Core Kit plus Licensed Probe, or license rights that may be purchased from Applied Biosystems. This product is an Authorized Core Kit without Licensed Probe. Its purchase price includes a limited, non-transferable immunity from suit under U.S. Patents and corresponding patent claims outside the United States, owned by Roche Molecular Systems, Inc. or F. Hoffmann-La Roche Ltd ("Roche"), for using only this amount of the product in the practice of the 5' nuclease process solely for the purchaser's own internal research and development activities. This product is also an Authorized Core Kit for use with service sublicenses available from Applied Biosystems. This product conveys no rights under U.S. Patents Nos. 5,804,375, 6,214,979, 5,538,848, 5,723,591, 5,876,930, 6,030,787, or 6,258,569, or corresponding patent claims outside the United States, expressly, by implication, or by estoppel. No right under any other patent claims (such as apparatus or system claims) and no right to perform commercial services of any kind, including without limitation reporting the results of purchaser's activities for a fee or other commercial consideration, is hereby granted expressly, by implication, or by estoppel. This product is for research purposes only. Diagnostic uses require a separate license from Roche. Further information regarding the 5' nuclease licensing program may be obtained from the Director of Licensing, Applied Biosystems, 850 Lincoln Centre Drive, Foster City, California 94404, USA.

---

## References

---

- Chou, Q., Russell, M., Birch, D., Raymond, J., and Bloch, W. (1992) Prevention of pre-PCR mis-priming and primer dimerization improves low-copy-number amplifications. *Nucl. Acids Res.*, 20, 1717-1723
- Lindahl, T., Ljungquist, S., Siegert, W., Nyberg, B., and Sperens, B. (1977) DNA N-glycosidases: properties of uracil-DNA glycosidase from *Escherichia coli*. *J. Biol. Chem.*, 252, 3286-3294
- Longo, M., Berninger, M., and Hartley, J. (1990) Use of uracil DNA glycosylase to control carry-over contamination in polymerase chain reactions. *Gene*, 93, 125-128
- Sharkey, D. J., Scalice, E. R., Christy, K. G., Atwood, S. M., and Daiss, J. L. (1994) Antibodies as thermolabile switches: high temperature triggering for the polymerase chain reaction. *Biotechnology*, 12, 506-509

---

©2010 Life Technologies Corporation. All rights reserved.

For research use only. Not intended for any animal or human therapeutic or diagnostic use. The trademarks mentioned herein are the property of Life Technologies Corporation or their respective owners.

# Notes

# Notes





**Corporate Headquarters**

5791 Van Allen Way  
Carlsbad, CA 92008

T: 1 760 603 7200

F: 1 760 602 6500

E: [tech\\_support@invitrogen.com](mailto:tech_support@invitrogen.com)

For country-specific contact information visit our web site at [www.invitrogen.com](http://www.invitrogen.com)