



ISO 9001 CERTIFIED
ISO 13485 CERTIFIED

Corporate Headquarters
AMRESCO, LLC
6681 Cochran Road
Solon, Ohio USA 44139-4300
440/349-1199 FAX: 440/349-1182
www.amresco-inc.com
Email: info@amresco-inc.com

CERTIFICATE OF QUALITY

Product: TE BUFFER pH 8.0
Grade: BIOTECHNOLOGY GRADE
Code: E112
Appearance: Colorless, haze-free liquid
Mol. Formula: Not Applicable
Mol. Weight: Not Applicable
Description: A blend of ACS Grade Tris Buffer and EDTA Disodium Salt, pH adjusted to 8.0.

<u>Assay:</u>	<u>Min</u>	<u>Max</u>	<u>Units</u>	<u>Disposition</u>
Conductivity @25C	775	975	umhos	PASS
DNase		NONE	P/F	PASS
pH @25C	7.90	8.10		PASS
Protease		NONE	P/F	PASS
Reassay Date	REPORT			PASS
RNase		NONE	P/F	PASS

Storage: Store at room temperature.

E112BIO

This signature indicates that the above material has met all quality specifications and has been reviewed by a quality representative.

Signature:

Title:

QC MANAGER

Date:

7/10/2012