



ISO 9001 CERTIFIED  
ISO 13485 CERTIFIED

Corporate Headquarters  
AMRESKO, LLC  
6681 Cochran Road  
Solon, Ohio USA 44139-4300  
440/349-1199 FAX: 440/349-1182  
www.amresco-inc.com  
Email: info@amresco-inc.com

CERTIFICATE OF QUALITY

**Product:** TRIS 1.0M pH 7.5 STERILE  
**Grade:** ULTRA PURE GRADE  
**Code:** E691  
Appearance: Clear, haze-free solution  
Mol. Formula: Not Applicable  
Mol. Weight: Not Applicable  
Description: A pH adjusted 1.0M Tris Buffer solution in specially prepared water (pH=7.5).

| <u>Assay:</u> | <u>Min</u> | <u>Max</u> | <u>Units</u> | <u>Disposition</u> |
|---------------|------------|------------|--------------|--------------------|
| DNase         |            | NONE       | P/F          | PASS               |
| pH@25C        | 7.4        | 7.6        |              | PASS               |
| Protease      |            | NONE       | P/F          | PASS               |
| RNase         |            | NONE       | P/F          | PASS               |
| Sterility     | PASS       |            | P/F          | PASS               |
| Tris          | 0.9        | 1.1        | M            | PASS               |

Storage: Store at room temperature. Protect from moisture.

*E691UP*

This signature indicates that the above material has met all quality specifications and has been reviewed by a quality representative.

Signature:

Title:

QC MANAGER

Date:

6/29/2012