


An Affymetrix Company

# Mouse IL-10 ELISA Ready-SET-Go!®

Catalog Number: 88-7104

Also known as: Interleukin-10, IL10

RUO: For Research Use Only. Not for use in diagnostic procedures.



Standard curve of Mouse IL-10 ELISA Ready-SET-Go!®

#### **Product Information**

Contents: Mouse IL-10 ELISA Ready-SET-

Go!®

REF Catalog Number: 88-7104 Sensitivity: 30 pg/ml

Standard Curve Range: 30 pg/mL - 4000

pg/mL



**Temperature Limitation:** Store at 2-8°C except standard which should be stored at less than or

equal to -70°C.



Batch Code: Refer to vial Use By: Refer to vial Caution, contains Azide



#### **Description**

This Mouse IL-10 ELISA Ready-SET-Go! reagent set (with or without high-affinity binding microwell plates) contains the necessary reagents, buffers and diluents for performing quantitative enzyme linked immunosorbent assays (ELISA). This ELISA reagent set is specifically engineered for accurate and precise measurement of mouse IL-10 protein levels from samples including serum, plasma and supernatants from cell cultures.

\*\*For best signal and lowest background, increase the stringency of washes for this assay. Be certain to follow the wash step as written in step 10, using 14 washes in total.

### Components

Capture Antibody. Pre-titrated, purified antibody

**Detection Antibody**. Pre-titrated, biotin-conjugated antibody

Standard. Recombinant cytokine for generating standard curve and calibrating samples

**ELISA/ELISPOT Coating Buffer Powder**. This Ready-Set-Go! ELISA Set may contain ELISA/ELISPOT Coating Buffer Powder (Reconstitute to 1L with dH20 and filter (0.22 uM)) or 10X PBS ELISA Coating Buffer (Dilute 1 part 10X Buffer into 9 parts dH20).

Assay Diluent. 5X concentrated

Detection enzyme. Pre-titrated Avidin-HRP

Substrate Solution. Tetramethylbenzidine (TMB) Substrate Solution

Certificate of Analysis. Lot-specific instructions for dilution of antibodies and standards

96 Well Plate. Corning Costar 9018 (included with product Cat. #'s ending in suffixes -22, -44, -76, -86)

#### References

Hsu SC, Tsai TH, Kawasaki H, Chen CH, Plunkett B, Lee RT, Lee YC, Huang SK. Antigen coupled with Lewis-x trisaccharides elicits potent immune responses in mice. J Allergy Clin Immunol. 2007 Jun;119(6):1522-8. (RSG ELISA, **TC sup**, PubMed)

Nakano T, Araki K, Nakatani H, Kobayashi M, Sugimoto T, Furuya Y, Matsuoka T, Jin T, Hanazaki K. Effects of



An Affymetrix Company

# Mouse IL-10 ELISA Ready-SET-Go!®

Catalog Number: 88-7104

Also known as: Interleukin-10, IL10

RUO: For Research Use Only. Not for use in diagnostic procedures.

geldanamycin and thalidomide on the Th1/Th2 cytokine balance in mice subjected to operative trauma. Surgery. 2007 Apr;141(4):490-500. (RSG ELISA, **serum**, PubMed)

Yang D, Zhang Y, Nguyen HG, Koupenova M, Chauhan AK, Makitalo M, Jones MR, St Hilaire C, Seldin DC, Toselli P, Lamperti E, Schreiber BM, Gavras H, Wagner DD, Ravid K. The A2B adenosine receptor protects against inflammation and excessive vascular adhesion. J Clin Invest. 2006 Jul;116(7):1913-23. (RSG ELISA, **plasma**, PubMed)

Mallen-St Clair J, Pham CT, Villalta SA, Caughey GH, Wolters PJ. Mast cell dipeptidyl peptidase I mediates survival from sepsis. J Clin Invest. 2004 Feb;113(4):628-34. (RSG ELISA, lavage fluid, PubMed)

#### **Related Products**

88-7044 Mouse IL-4 ELISA Ready-SET-Go!®

88-7121 Mouse IL-12 p70 ELISA Ready-SET-Go!®

88-7234 Mouse IL-23 ELISA Ready-SET-Go!® (Discontinued: Please see 88-7230 (2nd generation assay))

88-8314 Mouse IFNg 'Femto-HS' High Sensitivity ELISA Ready-Set-Go!



## **Research Use Only**

**Other Materials Needed** 

- Buffers
  - Wash Buffer: 1 x PBS, 0.05% Tween-20 (or eBioscience ELISA Wash Buffer Powder, cat 00-0400)
  - Stop Solution: 1M H<sub>3</sub>PO<sub>4</sub> or 2N H<sub>2</sub>SO<sub>4</sub>
- Pipettes and pipettors
- Refrigerator
- 96-well plate (Corning Costar 9018 or NUNC Maxisorp flat-bottom)

NOTE: The use of ELISA plates which are not high affinity protein binding plates will result in suboptimal performance, e.g., low signal or inconsistent data. Do not use tissue culture plates or low protein absorption plates. Use only the Corning Costar 9018 or NUNC Maxisorp 96 well plates provided or suggested.

- 96-well ELISA plate reader (microplate spectrophotometer)
- ELISA plate washer

NOTE: To ensure optimal results from this ELISA Ready-SET-Go! Set, please only use the components included in the set. Exchanging of components is not recommended as a change in signal may occur.

Stability

This ELISA set is guaranteed to perform as specified at least 6 months from date of receipt if stored and handled as instructed according to this datasheet and the Certificate of Analysis, which is included with the reagents.

**Storage Instructions for Cytokine Standards** 

The frozen cytokine standard is already aliquoted at 20 µl per vial. Upon receipt, frozen cytokine standard should be immediately stored at -80°C; stable for at least 6 months. After thawing, quick-spin vial prior to opening. Do not re-aliquot into smaller fractions. These are single use vials. Use one time and discard. For dilution of the standard, please see instructions on the Certificate of Analysis and follow these as written.

**Storage Instructions for Other Set Reagents** 

Store at 4°C.

**Time Requirements** 

- 1 overnight incubation
- 4½-hour incubations
- 1 hour washing and analyzing samples



## Research Use Only Experimental Procedure

NOTE: For best signal and lowest background, increase the stringency of washes for this assay. Be certain to follow the wash step as written in step 10.

- 1. Coat Corning Costar 9018 or NUNC Maxisorp 96 well ELISA plate with 100  $\mu$ l/well of capture antibody in Coating Buffer (dilute as noted on Certificate of Analysis, which is included with the reagent set). Seal the plate and incubate overnight at 4°C.
- 2. Aspirate wells and wash 5 times with >250  $\mu$ l/well Wash Buffer\*. Allowing time for soaking (~ 1 minute) during each wash step increases the effectiveness of the washes. Blot plate on absorbent paper to remove any residual buffer.
- 3. Dilute 1 part 5X concentrated Assay Diluent with 4 parts DI water\*. Block wells with 200 μl/well of 1X Assay Diluent. Incubate at room temperature for 1 hour.
- 4. Aspirate/wash as in step 2. Repeat for a total of 5 washes.
- 5. Using Assay Diluent\*, dilute standards as noted on the Certificate of Analysis (C of A). Add 100 μl/well of standard to the appropriate wells. Perform 2-fold serial dilutions of the top standards to make the standard curve. Add 100 μl/well of your samples to the appropriate wells. Cover or seal the plate and incubate at room temperature for 2 hours (or overnight at 4°C for maximal sensitivity).
- 6. Aspirate/wash as in step 2. Repeat for a total of 5 washes.
- 7. Add 100  $\mu$ l/well of detection antibody diluted in 1X Assay Diluent\* (dilute as noted on C of A). Seal the plate and incubate at room temperature for 1 hour.
- 8. Aspirate/wash as in step 2. Repeat for a total of 5 washes.
- 9. Add 100  $\mu$ l/well of Avidin-HRP\* diluted in 1X Assay Diluent (dilute as noted on C of A). Seal the plate and incubate at room temperature for 30 minutes.
- 10. Aspirate and wash as in step 2. In this wash step, soak wells in Wash Buffer\* for 1 to 2 minutes prior to aspiration. Repeat for a total of 14 washes.
- 11. Add 100 µl/well of Substrate Solution to each well. Incubate plate at room temperature for 15 minutes.
- 12. Add 50 μl of Stop Solution to each well.
- 13. Read plate at 450 nm. If wavelength subtraction is available, subtract the values of 570 nm from those of 450 nm and analyze data.

\*NOTE: Be certain that no sodium azide is present in the solutions used in this assay, as this inhibits HRP enzyme activity.

Ready-SET-Go Cytokine ELISA Set Buffers:

- Assay Diluent (5X concentrate): Dilute 1/5 in DI water.
- Substrate Solution: Ready to use (1X); 100 μl per well.
- ELISA/ELISPOT Coating Buffer Powder: Reconstitute in 1L dH20; filter (0.22 uM).



## **Research Use Only**

**Standard Calibration** 

The standard of the Ready-SET-Go! is calibrated against NIBSC standards:

Table of Standard Calibration					
Cytokine	ng of eB standard	ng of NIBSC standard	U of NIBSC standard	NIBSC Lot #	
hIL-2	1	1.1	14.6	86/564	
hIL-4	1	2.2	22	88/656	
hIL-5	1	2.2	22	90/586	
hIL-6	1	1.7	170	89/548	
hIL-10	1	0.8	4	93/722	
hIL-12	1	0.8	8	95/544	
hIFN-g	1	1.1	22	87/586	
hTNF-a	1	0.9	36	87/650	
mIL-2	1	3.1	310	93/566	
mIL-4	1	3	30	91/656	
mIL-6	1	8.5	850	93/730	
mIFN-g*	1		4.5	Gg02-901-533	
mTNF-a	1	1.7	340	88/532	
* Mouse IFN-g is calibrated using NIH standard (Lot Gg02-901-533) and is measured in Units (U)					

ELISA Troubleshooting Guide				
Problem	Possibility	Solution		
A. High Background	1. Improper and inefficient washing	1. Improve efficiency of washing. Fill plates completely, soak for 1 minute per wash, as directed		
	2. Cross contamination from other specimens or positive control	2. Repeat ELISA, be careful when washing and pipetting		
	3. Contaminated substrate	3. Substrate should be colorless		
	4. Incorrect dilutions, e.g., conjugate concentration was too high	4. Repeat test using correct dilutions; check with the recommendations of the antibody manufacturer		
B. No signal	1. Improper, low protein binding capacity plates were used	1. Repeat ELISA, using recommended high binding capacity plates		
	2. Wrong substrate was used	2. Repeat ELISA, use the correct substrate		
	3. Enzyme inhibitor present in buffers; e.g., sodium azide in the washing buffer and Assay Diluent inhibits peroxidase activity	3. Repeat ELISA, make sure your system contains no enzyme inhibitor		



# **Research Use Only**

Research ose only				
C. Very weak signal	1. Improper and inefficient washing	1. Make sure washing procedure is done correctly		
	2. Incorrect dilutions of standard	2. Follow recommendations of standard handling exactly as written on the certificate of analysis		
	3. Insufficient incubation time	3. Repeat ELISA, follow the protocol carefully for each step's incubation time		
	4. Incorrect storage of reagents	4. Store reagents in the correct temperature, avoid freeze and thaw, avoid using the "frost free" freezer		
	5. Wrong filter in ELISA reader was used	5. Use the correct wavelength setting		
	6. Wrong plate used	6. Use the recommended Corning Costar 9018 or NUNC Maxisorp flat bottom 96 well plates		
D. Variation amongst replicates	1. Improper and inefficient washing	Make sure washing procedure is done correctly; see certificate of analysis		
	2. Poor mixing of samples	2. Mix samples and reagents gently and equilibrate to proper temperature		
	3. Plates not clean	3. Plates should be wiped on bottom before measuring absorbance		
	4. Improper, low binding capacity plates were used	4. Use recommended high binding capacity plates		
	5. Reagents have expired	5. Do not use if past expiration date		