

It is highly recommended that this manual be read in its entirety before using this product. Do not use this kit beyond the expiration date.

For Research Purposes Only. Not for use in diagnostic or therapeutic procedures. Purchase does not include or carry the right to resell or transfer this product either as a stand-alone product or as a component of another product. Any use of this product other than the permitted use without the express written authorization of BioLegend is strictly prohibited.



Table of Contents

Page

Introduction	2
Materials Provided	2
Materials to be Provided by the End-User	3
Storage Information	3
Health Hazard Warnings	3
Specimen Collection and Handling	4
Reagent and Sample Preparation	4
Assay Procedure	5
Assay Procedure Summary	7
Calculation of Results	8
Typical Data	8
Performance Characteristics	9
Specificity	9
Sensitivity	9
Recovery	9
Linearity	9
Intra-Assay Precision	9
Inter-Assay Precision	9
Biological Samples	9
Troubleshooting Guide	10
ELISA Plate Template	12

Introduction:

Mouse IL-6 (other names include B-cell Stimulatory Factor 2, Interferon Beta-2, Hybridoma Growth Factor, and Cytotoxic T-Cell Differentiation Factor) plays an essential role in the final differentiation of B-cells into immunoglobinsecreting cells. It also induces myeloma and plasmacytoma growth, nerve cell differentiation, and acute phase reactant proteins in hepatocytes. IL-6 is expressed by T cells, B cells, monocytes, fibroblasts, hepatocytes, endothelial cells, and keratinocytes. Current research supports IL-6 as playing a critical role in many physiological and pathological conditions, including autoimmune diseases and rheumatoid arthritis.

The BioLegend LEGEND MAX[™] Mouse IL-6 ELISA Kit is a Sandwich Enzyme-Linked Immunosorbent Assay (ELISA) with a 96-well strip plate that is precoated with a capture antibody. This kit is specifically designed for the accurate quantitation of mouse IL-6 from cell culture supernatant, serum, plasma, and other biological fluids. This kit is analytically validated with ready-to-use reagents.

Description	Quantity (1 plate)	Quantity (5 plates)	Volume (per bottle)	Part #
Anti-mouse IL-6 Pre-coated 96-well Strip Microplate	1 plate	5 plates		78183
Mouse IL-6 Detection Anti- body	1 bottle	5 bottles	12 mL	78184
Mouse IL-6 Standard	1 vial	5 vials	lyophilized	78185
Avidin-HRP A	1 bottle	5 bottles	12 mL	79131
Assay Buffer A	1 bottle	5 bottles	25 mL	78232
Wash Buffer (20X)	1 bottle	5 bottles	50 mL	78233
Substrate Solution E	1 bottle	5 bottles	12 mL	78116
Stop Solution	1 bottle	5 bottles	12 mL	79133
Plate Sealers	4 sheets	20 sheets		78101

Materials Provided:

Materials to be Provided by the End-User:

- Microplate reader able to measure absorbance at 450 nm
- Adjustable pipettes to measure volumes ranging from 1 μL to 1,000 μL
- Deionized water
- Wash bottle or automated microplate washer
- Log-Log graph paper or software for data analysis
- Tubes to prepare standard dilutions
- Timer
- Plate Shaker
- Polypropylene vials

Storage Information:

Store unopened kit components at 4°C. Do not use this kit beyond its expiration date.

Opened or Reconstituted Components					
Microplate wells	If not all microplate strips are used, remove the excess strips by pressing up from underneath each strip. Place excess strips back in the foil pouch with the included des- iccant pack and reseal. Store at 4°C for up to one month.				
Standard	The remaining reconstituted standard stock solution can be aliquoted into polypropylene vials and stored at -70°C for up to one month. Avoid repeated freeze-thaw cycles.				
Detection Antibody					
Avidin-HRP A					
Assay Buffer A	Change and an angle of the and the state of the				
Wash Buffer (20X)	Store opened reagents at 4°C and use within one month.				
Substrate Solution E					
Stop Solution	1				

Health Hazard Warnings:

- 1. Reagents that contain preservatives may be harmful if ingested, inhaled or absorbed through the skin. Refer to the MSDS online at BioLegend's website for details (www.biolegend.com/msds).
- 2. Substrate Solution E is harmful if inhaled or ingested. Avoid skin, eye and clothing contact.

LEGEND MAX™ Mouse IL-6 ELISA Kit

- To reduce the likelihood of blood-borne transmission of infectious agents, handle all serum, plasma and other biological fluids in accordance with NCCLS regulations.
- 4. Stop Solution contains strong acid. *Wear eye, hand, and face protection.*
- 5. Before disposing of the plate, rinse it with an excess amount of tap water.

Specimen Collection and Handling:

Specimens should be clear and non-hemolyzed. If possible, unknown samples should be run at a number of dilutions to determine the optimal dilution factor that will ensure accurate quantitation.

<u>Cell Culture Supernatant</u>: If necessary, centrifuge all samples to remove debris prior to analysis. It is recommended that samples be stored at < -70°C. Avoid repeated freeze-thaw cycles.

<u>Serum</u>: Use a serum separator tube and allow clotting for at least 30 minutes, then centrifuge for 10 minutes at 1,000 x g. Remove serum layer and assay immediately or store serum samples at < -70°C. Avoid repeated freeze-thaw cycles.

<u>Plasma:</u> Collect blood samples in citrate, heparin or EDTA containing tubes. Centrifuge for 10 minutes at 1,000 x g within 30 minutes of collection. Assay immediately or store plasma samples at < -70°C. Avoid repeated freeze-thaw cycles.


Reagent and Sample Preparation:

Note: All reagents should be diluted immediately prior to use.

- Dilute the 20X Wash Buffer to 1X with deionized water. For example, make 1 liter of 1X Wash Buffer by adding 50 mL of 20X Wash Buffer to 950 mL of deionized water. If crystals have formed in the 20X Wash Buffer, bring to room temperature and vortex until dissolved.
- 2. Reconstitute the lyophilized Mouse IL-6 Standard by adding the volume of Assay Buffer A indicated on the vial label to make the 20 ng/mL standard stock solution. Allow the reconstituted standard to sit at room temperature for 15-20 minutes, then briefly vortex to mix completely.
- 3. In general, samples are analyzed without dilutions. However, if dilutions are required, use Assay Buffer A as the sample diluent.

Assay Procedure:


- Note: Do not mix reagents from different kits or lots. Reagents and/or antibodies from different manufacturers should not be used with this kit.
- 1. Bring all reagents to room temperature prior to use. It is strongly recommended that all standards and samples be run in duplicate or triplicate. A standard curve is required for each assay.
- 2. If not all microplate strips will be used, remove the excess strips by pressing up from underneath each strip. Place excess strips back in the foil pouch with the included desiccant pack and reseal.
- Prepare 500 µL of the 500 pg/mL top standard by diluting 12.5 µL of the standard stock solution in 487.5 µL of Assay Buffer A. Perform six two-fold serial dilutions of the 500 pg/mL top standard in separate tubes using Assay Buffer A as the diluent. Thus, the mouse IL-6 standard concentrations in the tubes are 500 pg/mL, 250 pg/mL, 125 pg/mL, 62.5 pg/mL, 31.3 pg/mL, 15.6 pg/mL and 7.8 pg/mL, respectively. Assay Buffer A serves as the zero standard (0 pg/mL).



- 4. Wash the plate 4 times with at least 300 μL of 1X Wash Buffer per well and blot any residual buffer by firmly tapping the plate upside down on absorbent paper. All subsequent washes should be performed similarly.
- 5. Add 50 μL of Assay Buffer A to each well that will contain either standard dilutions or samples.
- 6. Add 50 μ L of standard dilutions or samples to the appropriate wells
- 7. Seal the plate with a Plate Sealer included in the kit and incubate the plate at room temperature for 2 hours while shaking at 200 rpm.
- 8. Discard the contents of the plate into a sink, then wash the plate 4 times with 1X Wash Buffer as in step 4.
- 9. Add 100 μ L of Mouse IL-6 Detection Antibody solution to each well, seal the plate and incubate at room temperature for 1 hour while shaking.

- 10. Discard the contents of the plate into a sink, then wash the plate 4 times with 1X Wash Buffer as in step 4.
- 11. Add 100 μ L of Avidin-HRP A solution to each well, seal the plate and incubate at room temperature for 30 minutes while shaking.
- 12. Discard the contents of the plate into a sink, then wash the plate 5 times with 1X Wash Buffer as in step 4. For this final wash, soak wells in 1X Wash Buffer for 30 seconds to 1 minute for each wash. This will help minimize background.
- 13. Add 100 μ L of Substrate Solution E to each well and incubate for 15 minutes in the dark. Wells containing mouse IL-6 should turn blue in color with an intensity proportional to its concentration. It is not necessary to seal the plate during this step.
- 14. Stop the reaction by adding 100 μL of Stop Solution to each well. The solution color should change from blue to yellow.
- 15. Read absorbance at 450 nm within 30 minutes. If the reader is capable of reading at 570 nm, the absorbance at 570 nm can be subtracted from the absorbance at 450 nm.

Assay Procedure Summary



biolegend.com

LEGEND MAX™ Mouse IL-6 ELISA Kit


Calculation of Results:

The data can be best calculated with computer-based curve-fitting software using a 5- or 4-parameter logistics curve-fitting algorithm. If an appropriate software is not available, use log-log graph paper to determine sample concentrations. Determine the mean absorbance for each set of duplicate or triplicate standards, controls, and samples. Plot the standard curve on log-log graph paper with cytokine concentration on the X-axis and absorbance on the Y-axis. Draw a best fit line through the standard points. To determine the unknown cytokine concentrations, find the mean absorbance value of the unknown concentration on the Y-axis and draw a horizontal line to the standard curve. At the point of intersection, draw a vertical line to the X-axis and read the cytokine concentration.

If samples were diluted, multiply the concentration by the appropriate dilution factor. If a test sample's absorbance value falls outside the linear portion of the standard curve, the test sample needs to be re-analyzed at a higher (or lower) dilution as appropriate.

Typical Data:

This standard curve was generated at BioLegend for demonstration purposes only. A standard curve must be run with each assay.



Performance Characteristics:

<u>Specificity</u>: No cross-reactivity was observed when this kit was used to analyze mouse, rat, and human recombinant cytokines other than mouse IL-6.

Sensitivity: The minimum detectable concentration of IL-6 is 2 pg/mL.

<u>Recovery:</u> Recombinant IL-6 (200 and 50 pg/mL) was spiked into 4 mouse serum samples from 4 different strains, and then analyzed with the Legend Max[™] Mouse II-6 ELISA kit. On average, 92.2 % of the cytokine was recovered from the serum samples.

<u>Linearity</u>: Four murine serum samples, each from a diffrent strain, were spiked with high concentrations of IL-6 then diluted with Assay Buffer A to produce samples with concentrations within the dynamic range and then assayed. On average, 108% of the expected cytokine was detected from the serum samples.

<u>Intra-Assay Precision</u>: Sixteen replicates of each of two samples containing different IL-6 concentrations were tested in one assay.

Concentration	Sample 1	Sample 2
Number of Replicates	16	16
Mean Concentration (pg/mL)	97.0	24.6
Standard Deviation	5.5	1.4
% CV	5.7	5.6

<u>Inter-Assay Precision</u>: Two samples containing different concentrations of IL-6 were tested in four independent assays.

Concentration	Sample 1	Sample 2
Number of Assays	4	4
Mean Concentration (pg/mL)	101.8	24.8
Standard Deviation	10.8	1.4
% CV	10.7	5.8

<u>Biological Samples</u>: Mouse splenocytes (1×10^6 cells/mL) were stimulated with $1 \mu g/mL$ LPS overnight. Cell culture supernatant was collected and assayed for the concentrations of natural mouse IL-6. The IL-6 concentration averaged 36 pg/mL in LPS-stimulated samples and was undetectable in unstimulated samples.

Troubleshooting Guide:

Problem	Probable Cause	Solution		
High Background	Background wells were contaminated	Avoid cross-well contamination by using the provided plate sealers. Use multichannel pipettes and change tip between pipetting samples and reagents		
	Insufficient washes	Increase number of washes. Increase soaking time between washes prior to addition of substrate solution.		
	TMB Substrate Solution was contaminated	TMB Substrate Solution should be clear and colorless prior to addition to wells. Use a clean container prior to pipetting substrate solution into wells.		
No or poor signal	Detection Antibody, Avidin-HRP or Substrate solution were NOT added	Rerun the assay and follow the protocol.		
	Wrong reagent or reagents were added in wrong sequential order	Refult the assay and follow the protocol.		
	Insufficient plate agitation	The plate should be agitated during all incubation steps using a plate shaker at a speed where solutions in wells are within constant motion without splashing.		
	The wash buffer contains Sodium Azide (NaN3)	Avoid Sodium Azide contamination in the wash buffer as it inhibits HRP activity.		
	Incubations were done at an inappropriate temperature, timing or without agitation	Rerun the assay and follow the protocol.		
Low or poor standard curve signal	The standard was incorrectly reconstituted or diluted	Adjust the calculations and follow the protocol.		
	Standard was inappropriately stored	Store the reconstituted standard stock solution in polypropylene vials at -70°C. Avoid repeated freeze-thaw cycles.		
	Reagents added to wells with incorrect concentrations	Check for pipetting errors and the correct reagent volume.		

Problem	Probable Cause	Solution			
Signal is high, standard curves have saturated signal	Standard reconstituted with less volume than required	Reconstitute new lyophilized standard with the correct volume of solution recommended in the protocol.			
	Standards/samples, detection antibody, Avidin-HRP or substrate solution were incubated for too long	Rerun the assay and follow the protocol.			
Sample readings	Samples contain no or below detectable levels of the analyte	If samples are below detectable levels, it may be possible to use a larger sample volume. Contact technical support for appropriate protocol modifications.			
are out of range	Samples contain analyte concentrations greater than highest standard point	Samples may require dilution and analysis.			
	Multichannel pipette errors	Confirm that pipette calibrations are accurate.			
High variation in	Plate washing was not	Ensure pipette tips are tightly secured.			
samples and/or standards	adequate or uniform	Ensure uniformity in all wash steps.			
stanuarus	Non-homogenous samples	Thoroughly mix samples before assaying.			
	Samples may have high particulate matter	Remove particulate matter by centrifugation.			
	Cross-well contamination	Do not reuse plate sealers.			
		Always change tips for reagent additions. Ensure that pipette tips do not touch the reagents on the plate.			

	12								
	11								
	10								
	6								
a)	8								
ELISA Plate Template	7								
late Te	9								
ELISA P	5								
	4								
	3								
	2								
	1								
		А	В	С	D	E	F	Ð	Н

LEGEND MAX[™] Mouse IL-6 ELISA Kit **Notes**



LEGEND MAX[™] Kits are manufactured by **BioLegend Inc.** 9727 Pacific Heights Blvd. San Diego, CA 92121 Tel: 1.858.768.5800 Tel: US & Canada Toll-Free: 1.877.Bio-Legend (1.877.246.5343) Fax: 1.877.455.9587 Email: info@biolegend.com biolegend.com

For a complete list of world-wide BioLegend offices and distributors, please visit our website at: biolegend.com

78997-REV002-05022013