



Princeton Business Park, 5 Crescent Ave., P O Box 275
Rocky Hill, NJ 08553 USA
Tel: 609.497.0253 Fax: 609.497.0321
www.peprotech.com Email: info@peprotech.com

**Murine SCF
ELISA Development Kit
900-K78 Lot# 1107078
Expiration one year from date of receipt**

Description: Murine SCF ELISA development kit contains the key components required for the quantitative measurement of natural and/or recombinant mSCF in a sandwich ELISA format within the range of 23-2000pg/ml. Using the ELISA protocol described below, the recommended microplates, reagents and solutions, the components supplied in this kit are sufficient to assay mSCF in approximately 1000 ELISA plate wells.

RECONSTITUTION & STORAGE

Capture Antibody: 100µg of antigen-affinity purified rabbit anti-mSCF + 2.5mg D-mannitol. Centrifuge vial prior to opening. Reconstitute in 1ml sterile water for a concentration of 100µg/ml.

Detection Antibody: 100µg of biotinylated antigen-affinity purified rabbit anti-mSCF + 2.5mg D-mannitol. Centrifuge vial prior to opening. Reconstitute in 1ml sterile water for a concentration of 100µg/ml.

Murine SCF Standard: 1µg of recombinant mSCF + 2.2mg BSA + 11.0mg D-mannitol. Centrifuge vial prior to opening. Reconstitute in 1ml sterile water for a concentration of 1µg/ml.

Note: The reconstituted components are stable for 2 weeks when stored at 2-8°C. Components that have been reconstituted and aliquoted can be stored at -20°C for up to 6 months.

Avidin-HRP Conjugate: 60µl vial. Upon receipt, avidin-HRP conjugate should be aliquoted into ten 6µl vials and stored at ≤-20°C. Aliquots stored frozen at ≤-20°C are stable for up to 2 years form date of receipt. Avoid more than one freeze-thaw cycle. **Avidin should be used in conjunction with ABTS only.**

**RECOMMENDED MATERIALS (or purchase
PeproTech's ELISA Buffer Kit: Cat. # 900-K00)**

ELISA microplates (Nunc MaxiSorp Prod. # 439454, or Corning Prod # 3590);
Tween-20 (Sigma Cat. # P-7949);
BSA (Sigma Cat # A-7030);
ABTS Liquid Substrate Solution (Sigma Cat. # A3219);
Dulbecco's PBS [10x] (Gibco BRL Cat. # 14200-075).

RECOMMENDED SOLUTIONS

All solutions should be at ambient temperature prior to use.

PBS: dilute 10xPBS to 1xPBS, pH 7.20 in sterile water.

Wash Buffer: 0.05% Tween-20 in PBS

Block Buffer: 1% BSA in PBS *

Diluent: 0.05% Tween-20, 0.1% BSA in PBS *

*** Sterile filter and store at 4°C for up to 1 week.**

PLATE PREPARATION

1. Dilute capture antibody with PBS to a concentration of 1µg/ml. Immediately, add 100µl to each ELISA plate well. Seal the plate and incubate overnight at room temperature.

2. Aspirate the wells to remove liquid and wash the plate 4 times using 300µl of wash buffer per well. After the last wash invert plate to remove residual buffer and blot on paper towel.
3. Add 300µl block buffer to each well. Incubate for at least 1 hour at room temperature.
4. Aspirate and wash plate 4 times.

ELISA PROTOCOL

Standard/Sample: Dilute standard from 2ng/ml to zero in diluent. Immediately add 100µl of standard or sample to each well in triplicate. Incubate at room temperature for at least 2 hours.

Detection: Aspirate and wash plate 4 times. Dilute detection antibody in diluent to a concentration of 0.25µg/ml. Add 100µl per well. Incubate at room temperature for 2 hours.

Avidin-HRP Conjugate: Aspirate and wash plate 4 times. Dilute 5.5µl of avidin-HRP conjugate 1:2000 in diluent for a total volume of 11ml. Add 100µl per well. Incubate 30 minutes at room temperature.

ABTS Liquid Substrate:

(ABTS Substrate should be at ambient temperature prior to use)

Aspirate and wash plate 4 times. Add 100µl of substrate solution to each well. Incubate at room temperature for color development. Monitor color development with an ELISA plate reader at 405 nm with wavelength correction set at 650 nm.

NOTE: Reliable standard curves are obtained when either O.D. readings do not exceed 0.2 units for the zero standard concentrations, or 1.3 units for the highest standard concentration. The plate should be monitored at 5-minute intervals for approximately 35 minutes. *O.D. readings may vary.*

CROSS REACTIVITY

When tested at 50ng/ml the following antigens exhibited 100% cross reactivity:

Human	SCF
Rat	SCF

When tested at 50ng/ml the following antigens did not exhibit significant cross reactivity:

Human	flt3-L, GM-CSF, IL-1α, IL-1β, M-CSF
Murine	IL-1α, IL-1β, IL-4
Rat	IL-1α, IL-1β

