



ANPEP (Human) ELISA Kit

Catalog Number KA4261

96 assays

Version: 02

Intended for research use only

www.abnova.com

Table of Contents

Introduction	3
Intended Use	3
Background	3
Principle of the Assay	3
General Information	4
Materials Supplied	4
Storage Instruction	4
Materials Required but Not Supplied	4
Precautions for Use	5
Assay Protocol	6
Reagent Preparation	6
Sample Preparation	7
Assay Procedure	7
Data Analysis	9
Calculation of Results	9
Performance Characteristics	9
Resources	10
References	10
Plate Layout	11

Introduction

Intended Use

For quantitative detection of Human CD13 in cell culture supernates, serum and plasma (heparin, EDTA).

Background

Alanine aminopeptidase, also known as ANPEP or CD13, is an enzyme that is used as a biomarker to detect damage to the kidneys, and that may be used to help diagnose certain kidney disorders. It is mapped to 15q26.1. Aminopeptidase N is located in the small-intestinal and renal microvillar membrane, and also in other plasma membranes. In the small intestine, Aminopeptidase N plays a role in the final digestion of peptides generated from hydrolysis of proteins by gastric and pancreatic proteases, and it is also thought to be involved in the metabolism of regulatory peptides by diverse cell types, including small intestinal and renal tubular epithelial cells, macrophages, granulocytes, and synaptic membranes from the CNS. Petrovic et al showed that CD13 was required for endothelial cell invasion in response to bradykinin. Inhibition of CD13 abrogated internalization of bradykinin receptor B2 and reduced endothelial cell motility.

Principle of the Assay

ANPEP (Human) ELISA Kit was based on standard sandwich enzyme-linked immune-sorbent assay technology. A monoclonal antibody from mouse specific for CD13 has been precoated onto 96-well plates. Standards (NSO, K69-K967) and test samples are added to the wells, a biotinylated detection polyclonal antibody from goat specific for CD13 is added subsequently and then followed by washing with PBS or TBS buffer. Avidin-Biotin-Peroxidase Complex was added and unbound conjugates were washed away with PBS or TBS buffer. HRP substrate TMB was used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a blue color product that changed into yellow after adding acidic stop solution. The density of yellow is proportional to the human CD13 amount of sample captured in plate.

General Information

Materials Supplied

List of component

Component	Amount
96-well plate precoated with anti- Human CD13 antibody	96 (8x12) wells
Lyophilized recombinant human CD13 standard	20 ng/tube x 2
Biotinylated anti- human CD13 antibody (1:100)	130 μ L
Avidin-Biotin-Peroxidase Complex (ABC) (1:100)	130 μ L
Sample diluent buffer	30 mL
Antibody diluent buffer	12 mL
ABC diluent buffer	12 mL
TMB color developing agent	10 mL
TMB stop solution	10 mL

Storage Instruction

Store at 4 °C for 6 months, at -20 °C for 12 months. Avoid multiple freeze-thaw cycles.

Materials Required but Not Supplied

- ✓ Microplate reader in standard size.
- ✓ Automated plate washer.
- ✓ Adjustable pipettes and pipette tips. Multichannel pipettes are recommended in the condition of large amount of samples in the detection.
- ✓ Clean tubes and Eppendorf tubes.
- ✓ Washing buffer (neutral PBS or TBS).
 - Preparation of 0.01 M TBS:
Add 1.2 g Tris, 8.5 g NaCl; 450 μ L of purified acetic acid or 700 μ L of concentrated hydrochloric acid to 1000 mL H₂O and adjust pH to 7.2-7.6. Finally, adjust the total volume to 1 L.
 - Preparation of 0.01 M PBS:
Add 8.5 g sodium chloride, 1.4 g Na₂HPO₄ and 0.2 g NaH₂PO₄ to 1000 mL distilled water and adjust pH to 7.2-7.6. Finally, adjust the total volume to 1 L.

Precautions for Use

- ✓ To inspect the validity of experiment operation and the appropriateness of sample dilution proportion, pilot experiment using standards and a small number of samples is recommended.
- ✓ The TMB Color Developing agent is colorless and transparent before using, contact us freely if it is not the case.
- ✓ Before using the Kit, spin tubes and bring down all components to the bottom of tubes.
- ✓ Duplicate well assay is recommended for both standard and sample testing.
- ✓ Don't let 96-well plate dry, for dry plate will inactivate active components on plate.
- ✓ Don't reuse tips and tubes to avoid cross contamination.
- ✓ Avoid using the reagents from different batches together.
- ✓ In order to avoid marginal effect of plate incubation due to temperature difference (reaction may be stronger in the marginal wells), it is suggested that the diluted ABC and TMB solution will be pre-warmed in 37 °C for 30 min before using.

Assay Protocol

Reagent Preparation

- ✓ Reconstitution of the human CD13 standard : CD13 standard solution should be prepared no more than 2 hours prior to the experiment. Two tubes of CD13 standard (20 ng per tube) are included in each kit. Use one tube for each experiment.
 - 20,000 pg/mL of human CD13 standard solution: Add 1 mL sample diluent buffer into one tube, keep the tube at room temperature for 10 min and mix thoroughly.
 - 10,000 pg/mL→312 pg/mL of human CD13 standard solutions: Label 6 Eppendorf tubes with 10,000 pg/mL, 5000 pg/mL, 2500 pg/mL, 1250 pg/mL, 625 pg/mL, 312 pg/mL respectively. Aliquot 0.3 mL of the sample diluent buffer into each tube. Add 0.3 mL of the above 20,000 pg/mL CD13 standard solution into 1st tube and mix. Transfer 0.3 mL from 1st tube to 2nd tube and mix. Transfer 0.3 mL from 2nd tube to 3rd tube and mix, and so on.

Note: The standard solutions are best used within 2 hours. The 20 ng/mL standard solution should be stored at 4 °C for up to 12 hours, or at -20 °C for up to 48 hours. Avoid repeated freeze-thaw cycles.
- ✓ Preparation of biotinylated anti-human CD13 antibody working solution: The solution should be prepared no more than 2 hours prior to the experiment.
 - The total volume should be: 0.1 mL/well x (the number of wells). (Allowing 0.1-0.2 mL more than total volume)
 - Biotinylated anti-human CD13 antibody should be diluted in 1:100 with the antibody diluent buffer and mixed thoroughly. (i.e. Add 1 µL Biotinylated anti-human CD13 antibody to 99 µL antibody diluent buffer.)
- ✓ Preparation of Avidin-Biotin-Peroxidase Complex (ABC) working solution: The solution should be prepared no more than 1 hour prior to the experiment.
 - The total volume should be: 0.1 mL/well x (the number of wells). (Allowing 0.1-0.2 mL more than total volume)
 - Avidin- Biotin-Peroxidase Complex (ABC) should be diluted in 1:100 with the ABC dilution buffer and mixed thoroughly. (i.e. Add 1 µL ABC to 99 µL ABC diluent buffer.)

Sample Preparation

✓ Sample Preparation and Storage

Store samples to be assayed within 24 hours at 2-8 °C. For long-term storage, aliquot and freeze samples at -20 °C. Avoid repeated freeze-thaw cycles.

- Cell culture supernate: Remove particulates by centrifugation, analyze immediately or aliquot and store at -20 °C
- Serum: Allow the serum to clot in a serum separator tube (about 4 hours) at room temperature. Centrifuge at approximately 1000 x g for 15 min. Analyze the serum immediately or aliquot and store frozen at -20 °C.
- Plasma: Collect plasma using heparin or EDTA as an anticoagulant. Centrifuge for 10 min at 1000 x g within 30 min of collection. Analyze immediately or aliquot and store frozen at -20 °C.

✓ Sample Dilution Guideline

The user needs to estimate the concentration of the target protein in the sample and select a proper dilution factor so that the diluted target protein concentration falls near the middle of the linear regime in the standard curve. Dilute the sample using the provided diluent buffer. The following is a guideline for sample dilution. Several trials may be necessary in practice. The sample must be well mixed with the diluents buffer.

- High target protein concentration (200-2000 ng/mL). The working dilution is 1:100. i.e. Add 1 µL sample into 99 µL sample diluent buffer.
- Medium target protein concentration (20-200 ng/mL). The working dilution is 1:10. i.e. Add 10 µL sample into 90 µL sample diluent buffer.
- Low target protein concentration (312-20,000 pg/mL). The working dilution is 1:2. i.e. Add 50 µL sample to 50 µL sample diluent buffer.
- Very Low target protein concentration (\leq 312 pg/mL). No dilution necessary, or the working dilution is 1:2.

Assay Procedure

The ABC working solution and TMB color developing agent must be kept warm at 37 °C for 30 min before use. When diluting samples and reagents, they must be mixed completely and evenly. Standard CD13 detection curve should be prepared for each experiment. The user will decide sample dilution fold by crude estimation of CD13 amount in samples.

1. Aliquot 0.1 mL per well of the 20,000 pg/mL, 10,000 pg/mL, 5000 pg/mL, 2500 pg/mL, 1250 pg/mL, 625 pg/mL, 312 pg/mL human CD13 standard solutions into the precoated 96-well plate. Add 0.1 mL of the sample diluent buffer into the control well (Zero well). Add 0.1 mL of each properly diluted sample of human cell culture supernates, serum or plasma (heparin, EDTA) to each empty well. See "Sample Dilution Guideline" above for details. It is recommended that each human CD13 standard solution and each sample be measured in duplicate.

2. Seal the plate with the cover and incubate at 37 °C for 90 min.
3. Remove the cover, discard plate content, and blot the plate onto paper towels or other absorbent material. Do NOT let the wells completely dry at any time.
4. Add 0.1 mL of biotinylated anti-human CD13 antibody working solution into each well and incubate the plate at 37 °C for 60 min.
5. Wash plate 3 times with 0.01 M TBS or 0.01 M PBS, and each time let washing buffer stay in the wells for 1 min. Discard the washing buffer and blot the plate onto paper towels or other absorbent material. (Plate Washing Method: Discard the solution in the plate without touching the side walls. Blot the plate onto paper towels or other absorbent material. Soak each well with at least 0.3 mL PBS or TBS buffer for 1~2 minutes. Repeat this process two additional times for a total of THREE washes. *Note: For automated washing, aspirate all wells and wash THREE times with PBS or TBS buffer, overfilling wells with PBS or TBS buffer. Blot the plate onto paper towels or other absorbent material.*)
6. Add 0.1 mL of prepared ABC working solution into each well and incubate the plate at 37 °C for 30 min.
7. Wash plate 5 times with 0.01 M TBS or 0.01 M PBS, and each time let washing buffer stay in the wells for 1-2 min. Discard the washing buffer and blot the plate onto paper towels or other absorbent material. (See Step 5 for plate washing method).
8. Add 90 µL of prepared TMB color developing agent into each well and incubate plate at 37 °C in dark for 15-20 min (*Note: For reference only, the optimal incubation time should be determined by end user. And the shades of blue can be seen in the wells with the four most concentrated human CD13 standard solutions; the other wells show no obvious color*).
9. Add 0.1 mL of prepared TMB stop solution into each well. The color changes into yellow immediately.
10. Read the O.D. absorbance at 450 nm in a microplate reader within 30 min after adding the stop solution.

✓ Summary

1. Add samples and standards and incubate the plate at 37 °C for 90 min. Do not wash.
2. Add biotinylated antibodies and incubate the plate at 37 °C for 60 min. Wash plate 3 times with 0.01 M TBS.
3. Add ABC working solution and incubate the plate at 37 °C for 30 min. Wash plate 5 times with 0.01 M TBS.
4. Add TMB color developing agent and incubate the plate at 37 °C in dark for 15-20 min.
5. Add TMB stop solution and read.

Data Analysis

Calculation of Results

For calculation, (the relative O.D.₄₅₀) = (the O.D.₄₅₀ of each well) – (the O.D.₄₅₀ of Zero well). The standard curve can be plotted as the relative O.D.₄₅₀ of each standard solution (Y) vs. the respective concentration of the standard solution (X). The human CD13 concentration of the samples can be interpolated from the standard curve. *Note: if the samples measured were diluted, multiply the dilution factor to the concentrations from interpolation to obtain the concentration before dilution.*

✓ Typical Data Obtained from Human CD13

Concentration(pg/mL)	0	312	625	1250	2500	5000	10,000	20,000
O.D	0.067	0.144	0.213	0.444	0.720	1.300	1.811	2.463

(TMB reagent incubate at 37 °C for 15 min)

Performance Characteristics

- ✓ Range: 312 pg/mL-20,000 pg/mL
- ✓ Sensitivity: < 10 pg/mL
- ✓ Specificity: Natural and recombinant human CD13
- ✓ Cross-reactivity: No detectable cross-reactivity with other relevant proteins.
- ✓ Precision
 - Intra-Assay Precision (Precision within an assay) Three samples of known concentration were tested on one plate to assess intra-assay precision.
 - Inter-Assay Precision (Precision between assays) Three samples of known concentration were tested in separate assays to assess inter-assay precision.

Sample	Intra-Assay Precision			Inter-Assay Precision		
	1	2	3	1	2	3
n	16	16	16	24	24	24
Mean (ng/mL)	2.98	6.54	12.73	3.27	6.83	13.76
Standard deviation	0.158	0.360	0.751	0.190	0.430	0.922
CV (%)	5.3	5.5	5.9	5.8	6.3	6.7

Resources

References

1. "Entrez Gene: ANPEP alanyl (membrane) aminopeptidase "
2. Petrovic, N., Schacke, W., Gahagan, J. R., O'Connor, C. A., Winnicka, B., Conway, R. E., Mina-Osorio, P., Shapiro, L. H. CD13/APN regulates endothelial invasion and filopodia formation. *Blood* 110: 142-150, 2007.

Plate Layout

12								
11								
10								
9								
8								
7								
6								
5								
4								
3								
2								
1								
	A	B	C	D	E	F	G	H